



MISSOURI DEPARTMENT OF HEALTH  
STATE PUBLIC HEALTH LABORATORY  
DATAMASTER MAINTENANCE REPORT

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN <u>204139</u>	DATE OF INSPECTION <u>04-29-10</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>107 SOUTH MAIN, OREGON</u>	TIME OF INSPECTION <u>0937</u>

**CHECKLIST:** Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2 °C) 34.10 °C

☒ CALIBRATION CHECK -  
Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE  
(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 <input checked="" type="checkbox"/> <u>.101</u>	TEST 2 <input checked="" type="checkbox"/> <u>.101</u>	TEST 3 <input checked="" type="checkbox"/> <u>.102</u>
--	--	--

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS <input type="checkbox"/>	(0-.04) <input type="checkbox"/>	(.05-.09) <u>3</u>	(.10-.14) <input type="checkbox"/>	(.15-.19) <input type="checkbox"/>	(Over .19) <input type="checkbox"/>
-----------------------------------	----------------------------------	--------------------	------------------------------------	------------------------------------	-------------------------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

"OPERATING WITHIN MODEL'S SPECIFICATIONS"

REPCO MARKETING INC. # 09002, 08-31-11

INSPECTING OFFICER	
SIGNATURE <u>Joshua L. White</u>	PRINT NAME <u>JOSHUA L. WHITE</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>920113 / 05-11-11</u>	TELEPHONE NUMBER <u>(816) 387-2345</u>

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09002

EXPIRATION DATE: August 31, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is September 1, 2009.  
The expiration date for this lot number is August 31, 2011  
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marking, Inc.

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



JOSHUA L WHITE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 05/11/09

Number 920113

Expires 05/11/2011

MO 580-0771 (7-88)

*John J Mathewson*

Director of State Public Health Laboratory

*Margaret T. Donnelly*

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

## BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 2004139  
04/23/10  
09:42

--- SUPERVISOR NAME ---

COMPUTER: UNIT

PROGRAM (99 87 120000) 0000

BLANK  
SAMPLE CHAMBER 400

FLOW DETECTOR UNIT

PUMP UNIT

HIGH SPEED UNIT

DETECTOR UNIT

FLUORESC UNIT

CHAMBER CHAMBER UNIT

CHLORINE UNIT

PRINTED BY

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 2004139  
04/23/10  
09:42

Operator Signature

*TRR J.R. White*

Printed on recycled paper with agri-based inks

CMSU

Face This Side Down - This Edge In First

## BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 2004139  
04/23/10

TESTING OFFICER

WHITE/JOSHUA

OFFICER I.D.# 1200

PERMIT NUMBER: 920113

EXPIRATION DATE: 05/11/11

MISCELLANEOUS DATA

MAINTENANCE

--- SUPERVISOR NAME ---

BLANK TEST 1000 09:42

INTERNAL STANDARD VERIFIED 09:42

EXTERNAL STANDARD 110 09:42

BLANK TEST 1000 09:43

EXTERNAL STANDARD 110 09:43

BLANK TEST 1000 09:44

EXTERNAL STANDARD 110 09:44

BLANK TEST 1000 09:45

1000

1000

1000

Operator Signature

*TRR J.R. White*

Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204139  
04/29/10

ARREST TIME: 01:23  
SUBJECT NAME:  
X/X/X  
DOB: 01/23/45 SEX: M  
STATE/D.L.: MO/XXXXXXXXXX  
ARRESTING OFFICER:  
SAME  
OFFICER I.D.: SAME  
TESTING OFFICER:  
WHITE/JOSHUA/L  
OFFICER I.D.: 1202  
PERMIT NUMBER: 900115  
EXPIRATION DATE: 05/11/11  
MISCELLANEOUS DATA:  
MAINTENANCE

## --- BREATH ANALYSIS ---

BLANK TEST	.000	09:48
INTERNAL STANDARD	VERIFIED	09:48
RADIO INTERFERENCE		

Operator Signature



Printed on recycled paper with agri-based inks

CMSU 2208-02